



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT – FOR AGENCY USE ONLY

PES: SRS2459
Lab Name: Vista Analytical Laboratory
Contract: NBIA/NBFF PFAS
SDG No.: NR
Lab File ID: NR
Level: Low
Analysis Method: USEPA Method 537 Modified
Scoring Method: Manual
Comments: Scored by APTIM Federal Services, LLC Personnel

EPA Sample No.: MW-SRS2459
Case No.: 1900812
Matrix: Water
Date Received: 04/22/2019
Date Extracted: 04/29/2019
Sample Wt./Vol. (g/mL): 250 mL
pH: NA
Mod. Ref. No.: NR

Report Date: 05/14/2019
Lab Code: VAL
SAS/Client No.: NA
Lab Sample ID: 1900812-07
Date Analyzed: 05/02/2019
Dilution factor: 1.4
Units: ug/L

CAS No.	Analyte	Laboratory Results		PES Evaluation		PES Acceptance Limits				PES Statistics
		Concentration	Q			Action Low	Warning Low	Warning High	Action High	Spike Conc.
1763-23-1	Perfluorooctanesulfonic Acid (T-PFOS)	0.277		Pass	Within Limits	0.05	0.10	0.30	0.35	0.20
335-67-1	Perfluorooctanoic Acid (T-PFOA)	0.622		Pass	Warning High	0.10	0.20	0.60	0.70	0.40
*****	END All Analytes	*****	****	*****	*****	*****	*****	*****	*****	*****

NA = Not Applicable NR = Not Reported NE = Not Evaluated NL = No Limit ND = Not Detected above the Reporting Limit



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT

PES: SRS2459 **EPA Sample No.:** MW-SRS2459 **Report Date:** 05/14/2019
Lab Name: Vista Analytical Laboratory **Case No.:** 1900812 **Lab Code:** VAL
Contract: NBIA/NBFF PFAS **Matrix:** Water **SAS/Client No.:** NA
SDG No.: NR **Date Received:** 04/22/2019 **Lab Sample ID:** 1900812-07
Lab File ID: NR **Date Extracted:** 04/29/2019 **Date Analyzed:** 05/02/2019
Level: Low **Sample Wt./Vol. (g/mL):** 250 mL **Dilution factor:** 1.4
Analysis Method: USEPA Method 537 Modified **pH:** NA **Units:** ug/L
Scoring Method: Manual **Mod. Ref. No.:** NR
Comments: Scored by APTIM Federal Services, LLC Personnel

CAS No.	Analyte	Laboratory Results		PES Evaluation	
		Concentration	Q		
1763-23-1	Perfluorooctanesulfonic Acid (T-PFOS)	0.277		Pass	Within Limits
335-67-1	Perfluorooctanoic Acid (T-PFOA)	0.622		Pass	Warning High
****	End All Analytes	*****		*****	*****

NE = Not Evaluated
NR = Not Reported
NA = Not Applicable